


## AFLA

for analysis of Aflatoxin total (AFLA) in Animal feed



Aflatoxin B1 (AFLA)

### Specification

Lot No.	0031962080120275
Matrix Type	Animal feed
Analyte	Aflatoxin total (AFLA)
Weight/Volume	50 g
Storage	-18 °C
Retest	06/2024

	Concentration $x_{PT}$ [ $\mu\text{g}/\text{kg}$ ]	data points n	satisfactory range $x_{PT} \pm 2 \sigma_{PT}$ [ $\mu\text{g}/\text{kg}$ ]	uncertainty* $2 u(x_{PT})$ [ $\mu\text{g}/\text{kg}$ ]
Aflatoxin B1	12,07	15	4,56 - 19,57	2,86
Aflatoxin B2	3,09	9	1,73 - 4,45	0,90
Aflatoxin G1	6,20	11	3,47 - 8,93	0,68
Aflatoxin G2	presence**	7	nd	nd
Aflatoxin total	20,32	12	8,63 - 32,00	2,88

\*Expanded uncertainty with  $k=2$  for approximately 95% level of confidence (PA/PH/OMCL(18)153R1 CORR Evaluation of Measurement Uncertainty - Annex 2.5).

\*\*It was not possible to assign a value or an uncertainty due to a bimodal distribution of data. Arithmetic mean was 2,26  $\mu\text{g}/\text{kg}$  for Aflatoxin G2.

## Comments

**aokin** reference matrix material is naturally contaminated and homogenized. Concentration of the analyte is determinate in a proficiency round. Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination such as HPLC/MS/MS, HPLC/FLD, HPLC/UV, LC/MS/MS, UHPLC/MS/MS.

The certified value and its uncertainty are traceable to the International System of Units (SI) as chemical mass fraction as  $\mu\text{g}/\text{kg}$ . The assigned value, the satisfactory range and the expanded uncertainty are given.

The minimum amount of sample to be used is 10 g.

## Calculation of the assigned value $x_{PT}$

The Assigned Value is the value attributed to a particular property of interlaboratory proficiency test (definition from ISO13528:2016).  $x_{PT}$  is derived from participants quantitative results obtained with confirmatory analysis. The procedure for determining is from the Algorithm A (ISO 13528:2016) or from the median.

The standard uncertainty is expanded by a factor  $f = 1,25$  and is calculated as:

$$u(x_{PT}) = f \frac{\sigma_{PT}}{\sqrt{n}} \text{ (uncertainty of the characterization)}$$

where:

- $\sigma_{PT}$  is the robust estimate of the participant standard deviation;
- $\sigma_{PT} = b \cdot x_{PT}$  where  $b$  is the relative robust estimate of the participant standard deviation
- $n$  is the number of participants used in calculating the robust assigned values.

The satisfactory range is calculated from the expanded ( $k=2$ ) standard deviation of the proficiency assessment:

The satisfactory range equals  $x_{PT} \pm 2 \sigma_{PT}$ .