


OTA

ImmunoAffinity Column for Ochratoxin A for sample preparation



Ochratoxin A (OTA)



Manual

Specification

Lot No.	004A0XXXXXXX071
Recommended loading	< 200 ng OTA A
Loading rate	≤ 1 mL/min
Elution rate	< 1 mL/min
Expiry date	Date
	Recovery of 200 ng
Test results	Lot data
Specification	Recovery > 80 %

Methodology Testing

Sample loading	Standard	200 ng Ochratoxin A
	Fluid	Phosphate buffer/Methanol 90:10 v/v
	Volume	10 mL
	Rate	~ 1 mL/min
Column washing	Fluid	Phosphate buffer/Methanol 90:10 v/v
	Volume	5 mL
	Rate	> 1 mL/min
Elution	Eluant	0,004 % Glacial Acetic Acid / Methanol
	Volume	4 mL
	Rate	~ 1 mL/min
	Incubation	3 min
Analysis	by FPIA	