



ZON DON

for analysis of Zearalenone (ZON)
and Deoxynivalenol (DON) in Corn



Zearalenone (ZON)



Deoxynivalenol (DON)

Specification

| | |
|---------------|---|
| Lot No. | 122051210212275 |
| Matrix Type | Corn |
| Analyte | Zearalenone (ZON), Deoxynivalenol (DON) |
| Weight/Volume | 50 g |
| Storage | -18 °C |
| Retest | 05/2026 |

| | Concentration x_{PT} [$\mu\text{g}/\text{kg}$] | data points n | satisfactory range $x_{PT} \pm 2 \sigma_{PT}$ [$\mu\text{g}/\text{kg}$] | uncertainty* $2 u(x_{PT})$ [$\mu\text{g}/\text{kg}$] |
|----------------|---|------------------|--|---|
| Deoxynivalenol | 702,14 | 12 | 421,28 - 983,00 | 96,54 |
| Zearalenone | 232,61 | 16 | 130,26 - 334,95 | 58,14 |

*Expanded uncertainty with $k=2$ for approximately 95% level of confidence (PA/PH/OMCL(18)153R1 CORR Evaluation of Measurement Uncertainty - Annex 2.5).

Comments

aokin reference matrix material is naturally contaminated and homogenized. Concentration of the analyte is determinate in a proficiency round. Unweighted mean value of the means of accepted sets of data, each set being obtained in a different laboratory and/or with a different method of determination such as HPLC-MS/MS, HPLC/DAD, HPLC/FLD, HPLC/UV, HPLC/FLD, LC-MS/MS, UHPLC-MS/MS.

The certified value and its uncertainty are traceable to the International System of Units (SI) as chemical mass fraction as $\mu\text{g}/\text{kg}$. The assigned value, the satisfactory range and the expanded uncertainty are given.

The minimum amount of sample to be used is 10 g.

Calculation of the assigned value x_{PT}

The Assigned Value is the value attributed to a particular property of interlaboratory proficiency test (definition from ISO13528:2016). x_{PT} is derived from participants quantitative results obtained with confirmatory analysis. The procedure for determining is from the Algorithm A (ISO 13528:2016) or from the median.

The standard uncertainty is expanded by a factor $f = 1,25$ and is calculated as:

$$u(x_{PT}) = f \cdot \sigma_{PT} / \sqrt{n} \quad (\text{uncertainty of the characterization})$$

where:

- σ_{PT} is the robust estimate of the participant standard deviation;
- $\sigma_{PT} = b \cdot x_{PT}$ where b is the relative robust estimate of the participant standard deviation
- n is the number of participants used in calculating the robust assigned values.

The satisfactory range is calculated from the expanded ($k=2$) standard deviation of the proficiency assessment:

The satisfactory range equals $x_{PT} \pm 2 \sigma_{PT}$.